



TRI Analytical Methods

Analyte	Analysis Method	Level of Accuracy	Level of Precision	Minimum Detection Limit	Maximum Detection Limit
Carbon Monoxide	Gas Chromatography with catalytic converter followed by flame ionization detector	5% or ± 0.5 ppm, whichever is greater	$\pm 3\%$	0.5 ppm	2,000 ppm (100% special request)
Carbon Dioxide	Gas Chromatography with catalytic converter followed by flame ionization detector	5% or ± 5 ppm, whichever is greater	$\pm 3\%$	20 ppm (1ppm special request)	50,000 ppm (100% special request)
Methane	Gas Chromatography with catalytic converter followed by flame ionization detector	5% or ± 1 ppm, whichever is greater	$\pm 3\%$	1 ppm	2,000 ppm (100% special request)
Total Gaseous Hydrocarbons	Gas Chromatography with flame ionization detector	5% or ± 1 ppm, whichever is greater	$\pm 5\%$	1 ppm	2,000 ppm (100% special request)
Oxygen	Gas Chromatography with thermal conductivity detector	$\pm 2\%$ of concentration or 0.5% absolute whichever is greater	$\pm 2\%$	0.5%	100%
Nitrogen	Gas Chromatography with thermal conductivity detector	$\pm 2\%$ of concentration or 0.5% absolute whichever is greater	$\pm 2\%$	0.5%	100%
Condensed Hydrocarbons (Oil Mist and Particulates)	Standard Gravimetric, with hexane extraction for oil mist if results are within 90% of specification	± 0.1 mg/m ³ or 10% whichever is greater	$\pm 1\%$	0.1 mg/m ³	Varies
Moisture Dewpoint	Color indicator tube with critical orifice to measure air flow	± 8 ppm at 24 ppm $\pm 30\%$	$\pm 30\%$	-90°F	20°F
Nitrogen Dioxide	Color indicator tube with critical orifice to measure air flow	$\pm 30\%$	$\pm 20\%$	0.01 ppm	1.0 ppm
Sulphur Dioxide	Color indicator tube with critical orifice to measure air flow	$\pm 30\%$	$\pm 20\%$	0.1 ppm	3 ppm
Halogenated Solvents (Freon TF, 111-Trichloroethane, and others)	Gas Chromatography with electron capture	$\pm 10\%$	$\pm 10\%$	0.01 ppm	10 ppm
Halogenated Hydrocarbons (Freon TF, 111-Trichloroethane, and others)	Gas Chromatography with electron capture	$\pm 10\%$	$\pm 10\%$	0.1 ppm	Varies

Dr. Edward Golla, Ph.D., CIH, Laboratory Director

