

Supplementary Instructions to TRI's Lab for NFPA99 Medical Gas Tests

Please complete the below information that is applicable to your required analytical test type to ensure the laboratory performs the correct types of analysis and your final report has the correct information.

The price charged for your sample analysis report is based on the choices that you make on this sheet. Pricing information is available from our sales or order desks.

Analysis Type Selection Menu Check Box means this analysis will be performed by the lab if a proper test sample is submitted. Choose only those needed. Each check mark increases the price of the analysis.

NFPA99 1999 Edition

TEST NAME REFERENCE	Check Box	TEST REQUIREMENT
1999 Piping Purge 4-3.4.1.3(e)	<input type="checkbox"/>	← Matter - 0.1 mg (no breakdown of oil and particulate reported)
1999 Piping Purity Test 4-3.4.1.3(f) (Difference of Source vs. Outlet)	<input type="checkbox"/>	← Dew Point - 5°C @50psi (9 °F@50psi)
	<input type="checkbox"/>	← Total (non methane) Hydrocarbons - 1ppm (except in O2 and N2O)
	OR	or
	<input type="checkbox"/>	← Total (non methane) Hydrocarbons in O2 or N2O at additional cost
	<input type="checkbox"/>	← Halogenated Hydrocarbons - 2 ppm
1999 Medical Gas Conc. 4-3.4.1.3(l)	<input type="checkbox"/>	← N2, O2, N2O, CO2 gases >99%
	OR	or
	<input type="checkbox"/>	← Medical Air 19.5 to 23.5% O2
1999 Medical Air Purity Test (compressor system) 4-3.4.1.3(j)	<input type="checkbox"/>	← Dew Point - 39°F(4°C) @ 50 psi
	<input type="checkbox"/>	Carbon Monoxide - 10 ppm
		Carbon Dioxide - 500 ppm
		Total Hydrocarbons (as methane) - 25 ppm
		Halogenated Hydrocarbons - 2 ppm

NFPA99 2002 Edition

TEST NAME REFERENCE	TEST REQUIREMENT	
2002 Piping Particulate Test 5.1.12.3.7 or 2002 Verifier Piping Particulate Test 5.3.12.3.7 or equivalent	<input type="checkbox"/> ← Matter - 1.0 mg (no breakdown of oil and particulate reported)	
2002 Piping Purity Test 5.1.12.3.8 or 2002 Verifier Piping Purity Test 5.3.12.3.8 or equivalent (Difference of Source vs. Outlet except for Water Vapor)	<input type="checkbox"/> ← Water Vapor - 500 ppm Total	
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons - 1ppm (except in O2 and N2O)	
	OR	or
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons in O2 or N2O at additional cost	
	<input type="checkbox"/> ← Halogenated Hydrocarbons - 2 ppm	
2002 Medical Gas Conc. 5.1.12.3.11 or 2002 Verifier Medical Gas Conc. 5.3.12.3.11 or equivalent	<input type="checkbox"/> ← N2, O2, N2O, CO2 gases >99%	
	OR	or
	<input type="checkbox"/> ← Medical Air 19.5 to 23.5% O2	
2002 Medical Air Purity Test (compressor system) 5.1.12.3.12	<input type="checkbox"/> ← Dew Point - 39°F(4°C) @ 50 psi	
	<input type="checkbox"/>	Carbon Monoxide - 10 ppm
		Carbon Dioxide - 500 ppm
		Total Hydrocarbons (as methane) - 25ppm
		Halogenated Hydrocarbons - 2 ppm

NFPA99 2005 Edition

TEST NAME REFERENCE	TEST REQUIREMENT	
2005 Piping Particulate Test 5.1.12.3.7 or 2005 Verifier Piping Particulate Test 5.3.12.3.6 or equivalent	<input type="checkbox"/> ← Matter - 1.0 mg (no breakdown of oil and particulate reported)	
2005 Piping Purity Test 5.1.12.3.8 (Difference of Source vs. Outlet except for Water Vapor)	<input type="checkbox"/> ← Water Vapor - 500 ppm Total	
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons - 5 ppm (except in O2 and N2O)	
	OR	or
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons in O2 or N2O at additional cost	
	<input type="checkbox"/> ← Halogenated Hydrocarbons - 5 ppm	
2005 Verifier Piping Purity Test 5.3.12.3.7 or equivalent (Difference of Source vs. Outlet except for Water Vapor)	<input type="checkbox"/> ← Water Vapor - 500 ppm Total	
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons - 1 ppm (except in O2 and N2O)	
	OR	or
	<input type="checkbox"/> ← Total (non methane) Hydrocarbons in O2 or N2O at additional cost	
	<input type="checkbox"/> ← Halogenated Hydrocarbons - 2 ppm	
2005 Medical Gas Conc. 5.1.12.3.11	<input type="checkbox"/> ← O2 97%, N2, N2O, CO2 gases 99%	
	OR	or
	<input type="checkbox"/> ← Medical Air 19.5 to 23.5% O2	
2005 Verifier Medical Gas Conc. 5.3.12.3.10	<input type="checkbox"/> ← O2 or N2O, 99%	
2005 Medical Air Purity Test (compressor system) 5.1.12.3.12	<input type="checkbox"/> ← Dew Point - 39°F(4°C) @ 50 psi	
	<input type="checkbox"/>	Carbon Monoxide - 10 ppm
		Carbon Dioxide - 500 ppm
		Total Hydrocarbons (as methane) - 25ppm
		Halogenated Hydrocarbons - 2 ppm