

Choose the correct side of the datasheet based on sampling inside or outside of fill containment

1801 Central Commerce Court, Bldg 2  
Round Rock, TX 78664  
(512) 263-0498 Fax (512) 263-7039  
Email: info@airtesting.com

**DSC94**  70-DS-C94 Rev 4  
Next Sample Due Approx:  
Last Report No.  
Last Sample Date

All **Highlighted** Blanks Must be Complete  
Need Help? Call (512) 263-0498

Lab Use Only	Order Number	Kit Number	Report Number
--------------	--------------	------------	---------------

**Contact Information**

Customer Number	Company Name
Contact	E-mail
Phone	Testing Frequency

Account Information is pre-populated from last air test. Change above information if incorrect

**System Information** **Air Sample Collected**

Type of Compressor/Make/Model	<input type="radio"/> No Air Purification Filter Change
Serial Number	<input type="radio"/> Before Air Purification Filter Change
Date/Time Sample Collected	<input type="radio"/> After Air Purification Filter Change

**Add'l Information for Report (Optional)**

*Rush Analysis – Same Day Analysis - Additional Fee – By marking this box, I understand that I am authorizing Same Day Analysis & Reporting Please provide PO or credit card for additional fee (Normal Turnaround is one business day)*

**Air Spec:**

**Oil Mist/Particulate Sample Data**

Use a Separate Filter for Each Source/Sample.  
Do not touch or contaminate filter pad or metal screen.

Filter No				
Orifice Size Used	<table border="1"> <tr> <td>Choose One</td> <td><input type="radio"/> 0.115</td> <td><input type="radio"/> 0.200</td> </tr> </table>	Choose One	<input type="radio"/> 0.115	<input type="radio"/> 0.200
Choose One	<input type="radio"/> 0.115	<input type="radio"/> 0.200		
Flow Time	Minutes			

SCFM Orifice Plate	7 or fewer SCFM .115 orifice using 430 psi pressure reading from compressor	8 or more SCFM .200 orifice using 1000psi pressure reading from compressor
Minutes Required	12	12

**Moisture (H<sub>2</sub>O) Sample Data**

Immediately after sample is taken, record actual stain seen on the tube in the blank provided and on the picture below.

H<sub>2</sub>O Tube Serial #

Did Stain Occur? **Choose One**  Yes or  No

Record Reddish-Brown Stain Here and on picture below.

Break Both Tips of Tube Carefully, Approximately 1/8" from ends

Stain Greater Than Seven (7) STOP

If you are testing air for SCBA systems and your moisture tube shows a reddish-brown stain greater than seven (7) after you completed sampling as instructed, your sample is OUT OF SPECIFICATION per NFPA 1989 requirements. Please call TRI about what to do next.



**Breathing Air Sample Gas Sample Data**

Sample Bottle No	TRI's sample bottle must remain in place for a Minimum of one minute with the white float away from the <b>WHITE</b> sampling cap. <b>Replace Black Shipping Cap When Done</b>
Flow Time with Bottle in Place	

**Odor Sample Data**

Pronounced Odor? **Choose One** Yes  No

"Specific measurement of odor in gaseous air is impractical. Air normally may have a slight odor. The presence of a pronounced odor should render the air unsatisfactory."  
– From Section 5.6 of ANSI/CGA G-7.1-2011

By signing this data sheet, I declare that the sample submitted to TRI was taken according to the instructions provided and was not tampered with to falsify air test results.

Printed Name of Person Taking Test

Signature

Date